

	製品名	型番	試験結果に関する情報(学術論文等)	血圧計・カフのメンテナンス情報	掲載情報に関する連絡先
55	医用上腕式血圧計 (腕帯巻きつけタイプ)	HBP-1300	<p>ANSI/AAMI/ISO81060-2:2013</p> <p>●Linghui Meng, Di Zhao, Yan Pan, Wenqing Ding, Qing Wei, Hua Li, Pingjin Gao and Jie Mi. Validation of Omron HBP-1300 professional blood pressure monitor based on auscultation in children and adults. <i>BMC Cardiovascular Disorders</i> 2016, 16:9</p> <p>BHS</p> <p>●Lamaan Abbud, Diane Nzelu, Mariam Salaria, Polly Kay and Nikos A. Kametas. Validation of the Omron HBP-1300 in pregnancy for medium-arm and large-arm circumferences according to the British Hypertension Society protocol. <i>Blood Pressure Monitoring</i> 2018, 23:277-280</p>	<p>耐用年数 5年(適切な点検が行われた場合に限る)</p>	http://www.healthcare.omron.co.jp/support/
56	医用上腕式血圧計 (腕帯巻きつけタイプ)	HEM-907	<p>ESH International protocol</p> <p>●Stefano Omboni, Isabella Riva, Alessia Giglio, Gianluca Caldara, Antonella Gropelli and Gianfranco Parati. Validation of the Omron M5-I, R5-I and HEM-907 automated blood pressure monitors in elderly individuals according to the International Protocol of the European Society of Hypertension. <i>Blood Pressure Monitoring</i> 2007, 12:233-242</p> <p>ESH International protocol</p> <p>●Mohamed A. El Assaad, Jirar A. Topouchian, Bernadette M. Darne and Roland G. Asmar. Validation of the Omron HEM-907 device for blood pressure measurement. <i>Blood Pressure Monitoring</i> 2002, 7:237-241</p> <p>AAMI</p> <p>●William B. White and Yusra Anis Anwar. Evaluation of the overall efficacy of the Omron office digital blood pressure HEM-907 monitor in adults. <i>Blood Pressure Monitoring</i> 2001, 6:107-110</p> <p>BHS</p> <p>●Gregory K. Davis, Lynne M. Roberts, Geroge J. Mangos and Mark A. Brown. Comparisons of auscultatory hybrid and automated sphygmomanometers with mercury sphygmomanometry in hypertensive and normotensive pregnant women: parallel validation studies. <i>Journal of Hypertension</i> 2015, 33: 499-506</p>	<p>保守点検を行った場合、5年または3万回いずれか早く到達した方</p>	http://www.healthcare.omron.co.jp/support/
57	上腕式血圧計 (腕帯巻きつけタイプ)	HEM-7130-HP	<p>ESH International protocol revision 2010</p> <p>●Hakuo Takahashi, Masamichi Yoshika and Toyohiko Yokoi. Validation of three automatic devices for the self-measurement of blood pressure according to the European Society of Hypertension International Protocol revision 2010: the Omron HEM-7130, HEM-7320F, and HEM-7500F. <i>Blood Pressure Monitoring</i> 2015, 20:92-97</p>	<p>耐用期間 本体(腕帯を除く):30,000回もしくは5年のいずれか早く到達した方 腕帯:10,000回もしくは1年のいずれか早く到達した方</p>	http://www.healthcare.omron.co.jp/support/
58	上腕式血圧計 (腕帯巻きつけタイプ)	HEM-9200	<p>ESH International protocol revision 2010</p> <p>●Hakuo Takahashi, Masamichi Yoshika and Toyohiko Yokoi. Validation of three automatic devices for the self-measurement of blood pressure according to the European Society of Hypertension International Protocol revision 2010: the Omron HEM-7130, HEM-7320F, and HEM-7500F. <i>Blood Pressure Monitoring</i> 2015, 20:92-97</p>	<p>耐用期間 本体(腕帯を除く):30,000回もしくは5年のいずれか早く到達した方 腕帯:10,000回もしくは1年のいずれか早く到達した方</p>	http://www.healthcare.omron.co.jp/support/